

Wrist Blood **Pressure Monitor**

① EN

Model RS3 Instruction Manual



5338830-7B

Introduction

Thank you for purchasing the OMRON RS3 Wrist Blood Pressure Monitor.

The OMRON RS3 is a compact and easy to use blood pressure monitor, operating on the oscillometric principle. It measures your blood pressure and pulse rate simply and quickly. For comfortable controlled inflation without the need of pressure pre-setting or re-inflation the device uses its advanced "IntelliSense" technology.

Intended Use

Intended Use This product is designed to measure the blood pressure and pulse rate of people within the range of the designated wrist cuff, following the instructions in this instruction manual. It is mainly designed for general household use. Please read the Important Safety Information in this instruction manual before using

the unit.

 \fbox Please read this instruction manual thoroughly before using the unit. Please keep for future reference.

For specific information about your own blood pressure, CONSULT YOUR DOCTOR.

Important Safety Information

Consult your doctor prior to using in pregnancy or if diagnosed with arrhythmia or arteriosclerosis.

Please read this section carefully before using the unit.

 ${\rm I}{\Lambda}$ Warning: ${\rm \cdot}$ Indicates a potentially hazardous situation which, if not avoided, could result in death or serious injury.

- (General Usage) Always consult your doctor. Self-diagnosis of measurement results
- and self-treatment are dangerous. People with severe blood flow problems, or blood disorders, should consult a doctor before using the unit, as cuff inflation can cause internal bleeding.

(Battery Usage)

If battery fluid should get in your eyes, immediately rinse with plenty of clean water. Consult a doctor immediately.

▲ Caution: • Indicates a potentially hazardous situation which, if not avoided, may result in minor or moderate injury to the user or patient or damage to the equipment or other property.

- (General Usage) Do not leave the unit unattended with infants or persons who cannot express their consent.
- Do not use the unit for any purpose other than measuring blood pressure.
- Do not use a mobile phone or other devices that emit electromagnetic fields, near the unit. This may result in incorrect operation of the unit
- Do not disassemble the unit or wrist cuff. · Do not operate the unit in a moving vehicle (car, airplane).

(Battery Usage)

- If battery fluid should get on your skin or clothing, immediately rinse
- with plenty of clean water. Use only two "AAA" alkaline (LR03) batteries with this unit.
- Do not insert the batteries with their polarities incorrectly aligned
 Replace old batteries with new ones immediately. Replace both
- batteries at the same time. · Remove the batteries if the unit will not be used for three months or more.

· Do not use new and used batteries together.

General Precautions


- Do not apply strong shocks and vibrations to or drop the unit.
 Do not take measurements after bathing, drinking alcohol or
- caffeine, smoking, exercising or eating. Do not inflate the wrist cuff when it is not wrapped around your
- wrist.
- Read and follow the "Important information regarding Electro Magnetic Compatibility (EMC)" in the Technical Data Section.
- Read and follow the "Correct Disposal of This Product" in the Technical Data Section when disposing of the device and any used accessories or optional parts.

1. Overview





- A. Wrist cuff B. Battery compartment C. Display
- D. Memory (
) button
 E. Date/Time setting (
) button
 F. START/STOP button



Н.

Ι.

- K. Heartbeat symbol
- (Flashes during measurem L. Average value symbol

2. Preparation

2.1 Installing/Replacing the Batteries 1. Remove the battery cover.



Q. Cuff wrapping guide R. Movement error symbol S. Irregular heartbeat symbol

- 2. Insert two 1.5V "AAA" alkaline (LR03) batteries as indicated in the battery compartment and then replace the battery cover.


Notes: • If the battery low symbol ((a) appears on the display, turn off the unit then replace both batteries at the same time. • The measurement values continue to be stored in memory where the batteries are replaced.

even after the batteries are replaced.

Disposal of used batteries should be carried out in accordance with the national regulations for the disposal of batteries.

2.2 Setting the Date and Time

- 1. Press and hold the 🕒 button until the year flashes on the display.
- Set the monitor to the correct date and time 2. before taking a measurement for the first time.



- 3. Press the START/STOP button to store the setting. Notes:
 - If the batteries have been removed, the date and time setting will need to be reset.
 - If the date and time are not set. "-:--" appears during or after measurement

3. Using the Unit

3.1 Applying the Wrist Cuff

Do not apply over clothing.

1. Place the wrist cuff over your wrist.

Your palm should face upward

2. Wrap the wrist cuff around your wrist.

> Wrap the wrist cuff securely around the wrist for taking accurate measurements

Make sure that the wrist cuff does not cover the protruding part of the wrist bone (ulna) on the outside of the wrist.

Note: You can take a measurement on either your left or right wrist.

Download from Www.Somanuals.com. All Manuals Search And Download.

3.2 How to Sit Correctly

To take a measurement, you need to be relaxed and comfortably seated, under comfortable room temperature. No bathing, drinking alcohol or caffeine, smoking, exercising or eating 30 minutes before taking a measurement.

- Sit on a chair with your feet flat on the floor.
- Sit upright with your back straight.The cuff should be at the same level
- as your heart.


· Relax your wrist and hand. Do not bend your wrist back, clench your fist, or bend your wrist forward



3.3 Taking a Reading

Notes: • To cancel a measurement, press the START/STOP button at any time during measurement. Remain still while taking a measurement.

1. Press the START/STOP button. All the symbols appear on the display. The wrist cuff will start to inflate automatically.



Cuff Wrapping Guide The Cuff Wrapping Guide is a unique feature that indicates if the cuff is not wrapped tightly enough around the wrist. Even when the () is displayed, a blood pressure reading will be taken.

Note: This reading is **NOT** reliable due to the incorrect wrapping of the cuff. Please wrap the cuff again, taking care to wrap it correctly and take the measurement again. When the $\langle \mathbf{w} \rangle$ is displayed, the cuff is correctly wrapped tightly enough on the wrist and the reading is accurate and reliable.


2. Undo the wrist cuff and remove the unit.

3. Press the START/STOP button to turn off the monitor.

The monitor automatically stores the measurement in its memory

It will automatically turn off after two minutes

Important:





Diastolic Blood Pressure Above 85 mmHg This criteria is for home blood pressure measurement

 Your blood pressure monitor includes an irregular heartbeat feature. Irregular heartbeats can influence the results of the measurement. The irregular heartbeat algorithm automatically determines if the measurement is usable or needs to be repeated. If the measurement results are affected by irregular heartbeats but the result is valid, the result is shown together with the irregular heartbeat symbol (💟).

If you move during measurement, the

measurement.

A Warning:

Notes:

movement error symbol (
) will appear on the display. Keep still and repeat the

Wait 2 - 3 minutes before taking another blood pressure measurement. Waiting between readings allows the arteries to return to the condition prior to

The blood pressure measurement. The blood pressure can differ between the right arm and the left arm, and therefore also the measured

differ substantially, please check with your physician which arm to use for your measurement.

Self-diagnosis of measured results and treatment are dangerous. Please follow the instructions of your doctor.

blood pressure values can be different. OMRON recommends to always use the same arm for measurement. If the values between the two arms



1 18

18

ΠŌ

36

TTT I

If the irregular heartbeats cause the measurement to be invalid, no result is shown. If the irregular heartbeat symbol ()) is shown after you have taken a measurement, repeat the measurement. If the irregular heartbeat symbol ()) is shown frequently, please make your doctor aware of it.


3.4 Using the Memory Function

The unit automatically stores up to 60 sets of measurement values (blood pressure and pulse rate). It can also calculate an average reading based on the measurements from the last three readings taken within 10

minutes. If there are only two readings in memory for that period, the average will be based on two readings. If there is one reading in memory for that period, the average will be based on one reading. Notes:

- · If the memory is full, the monitor will delete the oldest
- readings.
- When viewing the reading taken without setting the date and time, "-:--" is displayed instead of the date and time.


To View the Average Value



To View the Readings Stored in Memory

1. Press the 🕑 button, while the average value is displayed.

The Memory number appears for a second before the pulse rate is displayed. The newest set is numbered "1".



2. Press the By button repeatedly to view the readings stored in memory

To Delete All the Values Stored in Memory

When the memory symbol () appears, first press the 🖹 button. Then while holding it down, press the START/STOP button simultaneously for about 2 - 3 seconds.



Note: You cannot partially delete the stored readings

4. Troubleshooting and Maintenance

4.1 The Icons and Error Messages

| Error Display | Cause | Remedy | |
|---------------|--|--|--|
| \bigcirc | Irregular heartbeats are detected. | Remove the wrist cuff. Wait 2 - 3 minutes and then take another measurement. Repeat the steps in section 3.3. If this error continues to appear, contact your doctor. | |
| 谷 | Movement during measurement. | Carefully read and repeat the steps in section 3.3. | |
| Blink | The batteries are low. | You should replace them with new ones ahead of time. Refer to section 2.1. | |
| 🔀 Lit | The batteries are exhausted. | You should replace them with new ones at once. Refer to section 2.1. | |
| E | Wrist cuff not applied correctly. | Apply the wrist cuff correctly. Refer to section 3.1. | |
| 63 | Movement during | Repeat Measurement while remaining still and refraining from talking during the | |
| E٩ | measurement. | measurement. Refer to section 3.3. | |
| 85 | Wrist cuff not applied correctly, or movement during measurement. | Apply the wrist cuff correctly and repeat measurement while remaining still and refraining from talking during the measurement. Refer to section 3.1 and 3.3. | |
| Er | Device error. | Contact your OMRON retail outlet or distributor. | |


| 4.2 Troubleshooting | 4.2 | ubleshooting |
|---------------------|-----|--------------|
|---------------------|-----|--------------|

| Problem | Cause | Remedy |
|--|--|--|
| | The wrist cuff is not at heart level. | Measure while in the correct posture. Refer to section 3.2. |
| The reading is extremely low | The cuff is not wrapped snugly around the wrist. | Wrap the cuff correctly. Refer to section 3.1. |
| (or high). | The arms and shoulders are tense. | Relax and try taking the measurement again. Refer to section 3.3. |
| | Movement or talking during measurement. | Remain still and do not talk during Measurement. Refer to section 3.3. |
| Wrist cuff pressure does not rise. | Air is leaking from the wrist cuff. | Consult your OMRON retail outlet or distributor. |
| Wrist cuff deflates too soon. | The wrist cuff is loose. | Apply the cuff correctly so that it is firmly wrapped around the wrist. Refer to section 3.1. |
| The blood pressure is different each time. The reading is extremely low (or high). | | Blood pressure readings constantly vary with time of day and how relaxed you are. Take several deep breaths and try to remain relaxed before taking a measurement. |
| The unit loses power during measurement. | The batteries are exhausted. | Replace the batteries with new ones. Refer to section 2.1. |
| Nothing happens when you press the buttons. | The batteries have been inserted incorrectly. | Insert the batteries with the correct (+/-) polarity. Refer to section 2.1. |
| Other problems. | measurement. If the problem con batteries with new If this still does not a | STOP button and repeat tinues, try replacing the ones. solve the problem, contact outlet or distributor. |

4.3 Maintenance

To protect your unit from damage, please observe the following

- Do not subject the main unit and the cuff to extreme temperatures, humidity, moisture or direct sunlight.
- Do not disassemble the unit.
- Do not subject the unit to strong shocks or vibrations (for example, dropping the unit on the floor).
- Do not use volatile liquids to clean the main unit.
- · Do not wash the cuff or immerse it in water.
- · Do not use petrol, thinners or similar solvents to clean the cuff.
- · Do not carry out repairs of any kind yourself. If a defect occurs, consult your OMRON retail outlet or distributor as mentioned on the packaging.



- . The unit should be cleaned with a soft, dry cloth. · Use a soft, moistened cloth and soap to clean
- the cuff. ·Keep the unit in its storage case when not in
- use
- · Fold the cuff into the storage case.
- Do not store the unit in the following situations:
 - If the unit is wet. · Locations exposed to extreme temperatures, humidity, direct sunlight, dust or corrosive vapours
 - · Locations exposed to vibrations, shocks or where it will be at an angle.

Calibration and Service

- The accuracy of this blood pressure monitor has been carefully tested and is designed for a long service life
- · It is generally recommended to have the unit inspected every two years to ensure correct functioning and accuracy. Please consult your authorised OMRON dealer or the OMRON Customer Service at the address given on the

5. Technical Data

Product Description Model Display Measurement Method Measurement Range

Inflation Deflation Memory Power Source Battery Life

Accuracy

Applied Part

Protection Against Electric Shock Operating temperature/ Humidity Storage temp Humidity Console Weight Outer Dimensions

Measurable circumference Cuff Material Package Content

Nylon and polyester Main unit, storage case, battery set, instruction manual, guarantee card

Approximately 13.5 to 21.5 cm

Wrist Blood Pressure Monitor

OMRON RS3 (HEM-6130-E)

Oscillometric method Pressure: 0 to 299 mmHg Pulse: 40 to 180/min.

Pulse: ±5% of display reading Automatic inflation by pump

2 "AAA" alkaline (LR03) batteries 1.5V

Approx. 300 measurements with new alkaline batteries at a room temperature of

Internally powered ME equipment

+10 to +40°C / Maximum: 30 to 85% RH

-20 to +60°C / Maximum: 10 to 95% RH /

Approximately 101g without batteries

Approximately 78 (w) mm x 60 (h) mm x 21 (d) mm (without the wrist cuff)

Automatic rapid deflation

LCD Digital Display

Pressure: ±3 mmHg

60 Measurements

700 to 1060hPa

23°C

Ň = Type B

Note: Subject to technical modification without prior notice

C€0197

- This device fulfils the provisions of EC directive 93/42/EEC (Medical Device Directive).
 This blood pressure monitor is designed according to the European Standard EN1060, Non-invasive sphygmomanometers Part 1: General Requirements and Part 3: Supplementary requirements
- for electromechanical blood pressure measuring systems. This OMRON product is produced under the strict quality system of OMRON HEALTHCARE Co. Ltd., Japan. The Core component for OMRON blood pressure monitors, which is the Pressure Sensor, is produced in Japan.

Important information regarding Electro Magnetic Compatibility (EMC)

With the increased number of electronic devices such as PC's and mobile (cellular) telephones, medical devices in use may be susceptible to electromagnetic interference from other devices Electromagnetic interference may result in incorrect operation of the medical device and create a potentially unsafe situation. Medical devices should also not interfere with other devices.

In order to regulate the requirements for EMC (Electro Magnetic Compatibility with the aim to prevent unsafe product situations, the EN60601-1-2:2007 standard has been implemented. This standard defines the levels of immunity to electromagnetic interferences as well as maximum levels of electromagnetic emissions for medical

This medical device manufactured by OMRON HEALTHCARE conforms to this EN60601-1-2:2007 standard for both immunity and emissions

Nevertheless, special precautions need to be observed: Do not use mobile (cellular) telephones and other devices, which generate strong electrical or electromagnetic fields, near the medical device. This may result in incorrect operation of the unit and create a potentially unsafe situation. Recommendation is to keep a minimum distance of 7 m. Verify correct operation of the device in case the distance is shorter.

Further documentation in accordance with EN60601-1-2:2007 is available at OMRON HEALTHCARE EUROPE at the address

mentioned in this instruction manual. Documentation is also available at <u>www.omron-healthcare.com</u>.

Correct Disposal of This Product (Waste Electrical & Electronic Equipment)

This marking shown on the product or its literature, indicates that it should not be disposed of, with other household wastes at the end of its working life.

To prevent possible harm to the environment or human health from uncontrolled waste disposal, please separate this product from other types of wastes and recycle it responsibly to promote the sustainable reuse of material resources.

Household users should contact either the retailer where they purchased this product, or their local government office, for details of where and how they can return this item for environmentally safe required. ecycling.

Business users should contact their supplier and check the terms and conditions of the purchase contract. This product should not be mixed with other commercial wastes for disposal.

| Manufacturer | OMRON HEALTHCARE Co., Ltd. |
|---------------------|--|
| | 53, Kunotsubo, Terado-cho, Muko, Kyoto, |
| | 617-0002 JAPAN |
| | |
| EU-representative | OMRON HEALTHCARE EUROPE B.V. |
| | Scorpius 33, 2132 LR Hoofddorp |
| EC REP | THE NETHERLANDS |
| | www.omron-healthcare.com |
| Production facility | OMRON (DALIAN) CO., LTD. |
| - | Dalian, CHINA |
| | |
| Subsidiary | OMRON HEALTHCARE UK LTD. |
| Subsidiary | Opal Drive, Fox Milne, Milton Keynes, MK15 0DG |
| Subsidiary | |
| Subsidiary | Opal Drive, Fox Milne, Milton Keynes, MK15 0DG |
| Subsidiary | Opal Drive, Fox Milne, Milton Keynes, MK15 0DG U.K. |
| Subsidiary | Opal Drive, Fox Milne, Milton Keynes, MK15 0DG U.K. OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH John-Deere-Str. 81a, 68163 Mannheim, GERMANY |
| Subsidiary | Opal Drive, Fox Milne, Milton Keynes, MK15 0DG U.K. OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH |
| Subsidiary | Opal Drive, Fox Milne, Milton Keynes, MK15 0DG U.K. OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH John-Deere-Str. 81a, 68163 Mannheim, GERMANY |
| Subsidiary | Opal Drive, Fox Milne, Milton Keynes, MK15 0DG U.K. OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH John-Deere-Str. 81a, 68163 Mannheim, GERMANY www.omron-medizintechnik.de |
| Subsidiary | Opal Drive, Fox Milne, Milton Keynes, MK15 0DG U.K. OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH John-Deere-Str. 81a, 68163 Mannheim, GERMANY www.omron-medizintechnik.de OMRON SANTÉ FRANCE SAS |

Made in China

Free Manuals Download Website <u>http://myh66.com</u> <u>http://usermanuals.us</u> <u>http://www.somanuals.com</u> <u>http://www.4manuals.cc</u> <u>http://www.4manuals.cc</u> <u>http://www.4manuals.cc</u> <u>http://www.4manuals.com</u> <u>http://www.404manual.com</u> <u>http://www.luxmanual.com</u> <u>http://aubethermostatmanual.com</u> Golf course search by state

http://golfingnear.com Email search by domain

http://emailbydomain.com Auto manuals search

http://auto.somanuals.com TV manuals search

http://tv.somanuals.com